

VERIFICATION OF CALIBRATION REPORT

Of Data Master cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statedwide Breath Alcohol Program

NOV 09 2009

DataMaster cdm S/N <u>130569</u> ✓	
Supervisor/Operator Performing the Verification Procedure:	
Name <u>John Hodges</u>	ID # <u>2999</u> Date <u>11/01/09</u> ✓
A	
Agency <u>Cordova Police Department</u>	Phone # <u>424-6100</u>
Instrument Location <u>Cordova Police Department</u>	
B	
Alco S/N <u>X301686</u>	Target Value <u>0.086</u> High Pressure <u>1100psi</u>
Alco Test Value Average <u>0.087</u>	<u>0.087</u>
1st Alco	2nd Alco
Signature <u>[Signature]</u>	
(OVER)	

DMB
11/20/09

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

(Do Not write in the area below.)


- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 21st day of Dec, 2009.

[Signature] (Notary Seal Stamp)

Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

Of Data Master cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statedwide Breath Alcohol Program

DataMaster cdm S/N

130569

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130569

NOVEMBER 01, 2009

OPERATOR'S NAME:

HODGES, JOHN

OPERATOR'S NUMBER: 2999

SUBJECT'S LAST NAME:

VERIFICATION OF CAL

SUBJECT'S FIRST NAME/MI :

X

O.L. #: X

DEPT/AGENCY: CDVI

FACE REPORT: X

TEST TYPE: V

ALCO TARGET VALUE: .086

ALCO S/N: X301606

--- BREATH ANALYSIS ---

.086 ADJUSTED FOR 29.78 in

ALCO TARGET .085 21:35

BLANK TEST .000 21:36

INTERNAL STANDARD VERIFIED 21:36

ALCO TV 29.78 in .087 21:36

BLANK TEST .000 21:37

SUBJECT SAMPLE .086 21:38

ALCO TEST .000 21:38

ALCO TV 29.78 in .087 21:39

BLANK TEST .000 21:40

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130569

NOVEMBER 02, 2009

TIME 01:15

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 02/28/01

HEATERS

SAMPLE CHAMBER: 49c

BREATH TUBE: 48c

BAROMETER: 29.77 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

11/02/09 01:15:15
ALASKA DEPARTMENT OF PUBLIC SAFETY
SCIENTIFIC CRIME DETECTION LABORATORY
STATEDWIDE BREATH ALCOHOL PROGRAM